



Namibia Institute of Pathology Limited

CLARIFICATION NO.: 01

PROCUREMENT TITLE: Supply and Delivery of Laboratory Equipment

CLARIFICATION QUESTIONS AND ANSWERS – Procurement Reference No: G/RFO/NIP-28/2023

Please find in the below table feedback to clarifications sought from bidder(s):

ITEM	ITEM	COMMENTS/ RESPONSE
1.	ITEM LINE NO 9 (AUTOMATED TEMPERATURE MONITORING SYSTEM), We would like to know the number of <b>sensors</b> required for the Fridges, Freezers, incubators and for the room temperature to monitor. Kindly provide us with that information.	<p>Sensors needed:</p> <ul style="list-style-type: none"><li>• 3 – for the fridge</li><li>• 3 – for the freezers</li><li>• 6 – for room temperatures</li></ul> <p>The laboratory does not have an incubator.</p>

Yours sincerely

Hilde Kautondokua  
ASSISTANT PROCUREMENT OFFICER  
NAMIBIA INSTITUTE OF PATHOLOGY LIMITED (NIP)

2023-11-17

